

Vitassay qPCR

Chikungunya

PCR en tiempo real para la detección cualitativa del virus Chikungunya en muestras humanas

Real-time PCR kit for the qualitative detection of Chikungunya virus in human samples



Uso previsto

Vitassay qPCR Chikungunya permite la detección cualitativa de Chikungunya virus mediante RT-PCR en tiempo real en muestras clínicas. Este producto está destinado para facilitar el diagnóstico de infecciones producidas por el virus Chikungunya.

Referencias

Vitassay qPCR Chikungunya 4 x 8-well strip, low profile	7041001
Vitassay qPCR Chikungunya 4 x 8-well strip, high profile	7042001

Materiales/Reactivos suministrados

Código	Reactivo/Material	Color	Cantidad
7041S001/ 7042S001	Chikungunya Virus strips high/low profile	-	4 tiras de 8 pocillos
7C001	Chikungunya Virus Positive Control	rojo	1 vial
7001A	PCR grade water	blanco	1 vial x 1 mL
7002B	Resuspension buffer	verde	1 vial x 1 mL
7003N	Negative control	amarillo	1 vial x 1 mL
7004O	Tapas ópticas	-	4 tiras de 8 tapones

Condiciones de Transporte y conservación.

- El transporte y almacenaje de los kits puede realizarse de 2-40°C hasta la fecha de caducidad indicada en la etiqueta.
- El control positivo resuspendido debe ser almacenado a -20°C. Para evitar ciclos de congelación y descongelación, se recomienda separar en alícuotas.
- Conservar los reactivos en la oscuridad.

Material y equipamiento necesario pero no proporcionado

- Kit de extracción de RNA
- Equipo de PCR en tiempo real (ver Adjunto I)
- Centrífuga para tubos de 1.5 mL
- Vórtex
- Micropipetas (1-20 µL, 20-200 µL)
- Puntas con filtro
- Guantes desechables sin polvo

Resumen

El virus Chikungunya (CHIKV) es un virus que se transmite a través de mosquitos (arbovirus) el cual se descubrió durante un brote en el sur de Tanzania en 1952. Se transmite a humanos principalmente a través de mosquitos infectados del género *Aedes*, en concreto *Ae. aegypti* y *Ae. albopictus*. Aunque la transmisión perinatal vertical también se ha observado.

Desde 1952, CHIKV ha causado numerosos brotes y epidemias bien documentadas en África, Asia, Europa, el Sur del Pacífico y las Américas, implicando a millones de personas y a menudo en periodos de más de 10 años. Tras la picadura de un mosquito infectado, el inicio de la enfermedad se produce entre 4 y 8 días, aunque puede variar de 2 a 12 días. La infección por Chikungunya se caracteriza por un inicio brusco con escalofríos, fiebre hasta alcanzar 40 °C, vómitos, náuseas, dolor de cabeza, artralgia (dolor articular) y en algunos casos erupción maculopapular. El síntoma principal y el más problemático del virus Chikungunya es el dolor articular y muscular severo. El dolor es tan intenso que dificulta el movimiento de los infectados y los postra.

El diagnóstico en laboratorio se realiza generalmente analizando suero o plasma para detectar virus, ácido nucleico viral o inmunoglobulina específica del virus (Ig M y anticuerpos neutralizantes). Durante los primeros 8 días de la enfermedad, el RNA viral de Chikungunya se puede detectar en suero. En cambio, los anticuerpos frente al virus Chikungunya se desarrollan por lo general al final de la primera semana de la enfermedad, por lo que para descartar la infección, es recomendable recoger también muestras durante la fase convaleciente en aquellos pacientes con un resultado negativo en la fase aguda.

La técnica de PCR a Tiempo Real es un método muy común de detección debido a que las reacciones cruzadas de los anticuerpos de Chikungunya con otros arbovirus dificultan la serología.

Principio del test

El test Vitassay qPCR Chikungunya está diseñado para la identificación de virus Chikungunya en muestras clínicas como ayuda en la evaluación de las infecciones producidas por este virus.

Vitassay qPCR Chikungunya se basa en la amplificación a tiempo real de un fragmento de una región conservada de la zona del gen *NSP1* del virus Chikungunya. El RNA viral

extraído se transcribe a cDNA utilizando un *primer* específico mediante un paso de transcripción inversa seguido, de la reacción en cadena de la polimerasa. La presencia de virus Chikungunya se detecta mediante un aumento de la fluorescencia observada durante la reacción, tras la hidrólisis de la sonda fluorescente.

El ensayo está basado en la actividad 5' nucleasa que utiliza dos *primers* y una sonda de hidrolisis fluorogénica para detectar la acumulación de la secuencia diana amplificada durante la reacción de PCR. Cuando la polimerasa comienza a extender los primers, la sonda es hidrolizada mediante su actividad exonucleasa 5'- 3' produciendo la separación espacial del fluoróforo y el *quencher*. El aumento de la señal fluorescente resultante es proporcional a la cantidad de producto amplificado en la muestra y es detectado mediante un equipo de PCR en tiempo real.

Vitassay qPCR Chikungunya se trata de un ensayo listo para usar que contiene en cada pocillo todos los reactivos necesarios en formato estabilizado para llevar a cabo la PCR en tiempo real. Además un control interno permite la detección de una posible reacción de inhibición. La amplificación de la secuencia diana es detectada en el canal FAM mientras que el control interno (CI) se detecta en el canal HEX, VIC o JOE (según el equipo utilizado).

Precauciones

- Diseñado para uso profesional de diagnóstico *in vitro*.
- No utilizar el kit después de la fecha de caducidad.
- No mezclar reactivos de otros kits y/o diferentes lotes.
- No utilizar si el kit tiene signos de haber sido abierto o manipulado.
- Trabajar siguiendo las Buenas Prácticas de Laboratorio. Use ropa protectora, guantes desechables, gafas y mascarilla.
- No comer, beber o fumar en los laboratorios.
- Es importante seguir un flujo de trabajo en el laboratorio unidireccional: Área de Extracción, Área de Amplificación y Detección. No retornar muestras, equipos ni reactivos a un área anterior.
- Las muestras y todo material en contacto con ellas se deben tratar como potencialmente infecciosos y se deben gestionar según la legislación sobre residuos sanitarios nacional. Tome las precauciones necesarias durante la recogida, almacenamiento, tratamiento y eliminación de muestras.
- Se recomienda la descontaminación periódica de los equipos usados habitualmente, especialmente micropipetas, y de las superficies de trabajo.

Procedimiento

Toma de muestra, preparación y extracción de RNA.

Realizar el pretratamiento y el aislamiento de los ácidos nucleicos utilizando un sistema manual o automático compatible con ensayos de PCR en tiempo real. Seguir las instrucciones de uso del fabricante. Los siguientes kits de extracción han sido validados:

Maxwell® 16 Viral Total Nucleic Acid Purification Kit (Promega).

QIAamp Viral RNA Mini Kit, QIAcube instrument (QIAGEN).

Invisorb® Spin Universal Kit (Stratec).

Preparación del control positivo

Reconstituir el Chikungunya Virus Positive Control (tubo rojo) liofilizado con 100 µL de PCR grade wáter suministrado. Mezclar bien con ayuda del vórtex. Después del primer uso, dispensar en alícuotas para evitar varios ciclos de congelación-descongelación y almacenarlo a -20°C.

El control positivo contiene una gran cantidad de copias molde y el riesgo de contaminación es elevado. Por lo tanto, se recomienda abrir y manipular en un área del laboratorio separada de los otros componentes.

Preparación de la reacción

- Preparar el número de pocillos necesarios incluyendo muestras y controles (un control positivo y uno negativo). Retirar el sello de aluminio que protege las tiras.
- Pipetear 15 µL de la solución de resuspensión (tubo verde) en cada pocillo.
- Pipetear 5 µL de RNA extraído, Control Negativo (tubo amarillo) y Control positivo (tubo rojo) en los pocillos correspondientes.
- Cerrar los pocillos con los tapones suministrados. Centrifugar brevemente (opcional).
- Colocar las tiras en el equipo de PCR en tiempo real.

Programación del termociclador.

Configurar el termociclador siguiendo las siguientes instrucciones:

Etapa	Temperatura	Tiempo	Ciclos
Retrotranscripción	45°C	15 min	1
Desnaturalización inicial	95°C	2 min	1
Desnaturalización	95°C	10 seg	45
Hibridación/Elongación (Recogida de datos*)	60°C	50 seg	

Los datos de fluorescencia deben recogerse durante la etapa de hibridación (*) a través de los canales FAM (virus Chikungunya) y los canales HEX, JOE o VIC (Control Interno). En los termocicladores Applied Biosystems 7500 Fast, Applied Biosystems StepOne™, y Stratagene Mx3005P™ comprobar que la opción del control pasivo ROX esta desactivada (ver adjunto II).

Análisis e interpretación de resultados

El análisis de las muestras se realiza con el software propio del equipo de PCR en tiempo real de acuerdo con las instrucciones de uso del fabricante.

Antes de analizar el resultado de las muestras clínicas debe validarse el resultado de los controles:

Control positivo El control positivo utilizado en cada serie, debe mostrar una curva de amplificación en el canal FAM.

Control negativo El control negativo incluido en cada serie, debe mostrar la ausencia de señal de FAM.

El experimento es inválido si hay señal de amplificación en el control negativo o ausencia de la señal en el control positivo. El ensayo se debe de repetir.

Con la ayuda de la siguiente tabla, analizar los resultados:

Virus Chikungunya FAM	Control Interno HEX	Control Negativo	Control Positivo	Interpretación
+	+/-	-	+	Virus Chikungunya Positivo
-	+	-	+	Virus Chikungunya Negativo
+	+	+	+	Inválido
-	-	-	-	Inválido

+ Positivo: Señal de amplificación

- Negativo: No hay señal de amplificación

Si las muestras negativas no muestran un resultado positivo para el control interno, se debe repetir el ensayo diluyendo la muestra original 1:10 o repetir la extracción de los ácidos nucleicos debido a posibles problemas causados por inhibidores de PCR.

Control de Calidad

Con el fin de confirmar el correcto funcionamiento de la técnica de diagnóstico molecular, se incluye un Control Interno (CI) en cada reacción. Además un control positivo y uno negativo se debe de incluir en cada ensayo para una correcta interpretación de los resultados.

Características técnicas

Sensibilidad y especificidad clínica

Un total de 10 muestras clínicas disueltas en un medio de transporte fueron evaluadas mediante PCR en tiempo real utilizando: Vitassay qPCR Chikungunya y RealStar® Chikungunya Virus RT-PCR Kit (Altona Diagnostics). El virus del Chikungunya fue detectado en todas las muestras mediante el test Vitassay qPCR Chikungunya. Los resultados fueron concordantes entre ambos test. Vitassay Chikungunya qPCR muestra una alta sensibilidad y especificidad para detectar el virus del Chikungunya.

Sensibilidad analítica

La sensibilidad analítica fue determinada a a partir de diluciones seriadas (1:10) del RNA molde de virus Chikungunya (10^7 - 10^1 copias/reacción). Este ensayo tiene un límite de detección de ≥ 10 copias de RNA viral por reacción.

Especificidad analítica

La especificidad analítica para la detección de virus Chikungunya fue confirmada probando un panel compuesto por los siguientes microorganismos, no observándose reacciones cruzadas entre ninguna de las especies:

Chikungunya Virus		
Virus Zika cepa MR 766	Virus Dengue 4 cepa H241	Virus West Nile Heja
Virus Dengue 1 cepa Hawaii	Virus de la encefalitis de San Luis cepa 17D	Virus West Nile Ug37
Virus Dengue 2 cepa New Guinea C	Virus West Nile cepa H160/99	Virus de la Fiebre Amarilla cepa 17D
Virus Dengue 3 cepa H87		

Reactividad analítica

La reactividad de Vitassay qPCR Chikungunya fue confirmada por medio de la amplificación a tiempo real utilizando Chikungunya cepa S27Petersfield como molde.

Termocicladores compatibles

Vitassay qPCR Chikungunya ha sido validado en los siguientes equipos:

- LightCycler Z480, 480II (Roche)
- Cobas Z480 (Roche)
- 7500 Fast Real-Time PCR System (Applied Biosystems)
- StepOne™ Real-Time PCR System (Applied Biosystems)
- CFX96 Touch™ Real-Time PCR Detection System (Bio-Rad)
- AriaMx Real-Time PCR System (Agilent Technologies)
- DTPRime Real Time Detection Thermal Cycler (DNA-Technology)
- DTLite Real Time PCR System (DNA-Technology)
- Rotor-Gene® Q (Qiagen)
- SmartCycler® (Cepheid)

Para los equipos Rotor-Gene® Q y SmartCycler® el producto debe ser reconstituido y trasvasado a los tubos específicos de cada uno de los equipos.

En el caso de utilizar el equipo Applied Biosystems 7500 Fast con tiras, se recomienda colocar el soporte adecuado para los tubos (Ref. PN 4388506).

Limitaciones

- Este test proporciona un diagnóstico preliminar de infección por virus Chikungunya. Todos los resultados obtenidos deben ser interpretados por un especialista junto con la información clínica y los hallazgos de laboratorio disponibles.
- Este ensayo ha sido probado con muestras clínicas disueltas en medio de transporte. El uso de otras muestras no se ha establecido.
- El correcto funcionamiento de la prueba depende de la calidad de la muestra; el RNA deber ser extraído de forma adecuada de las muestras clínicas humanas. Una forma inadecuada de recolección, almacenaje y/o transporte de las muestras puede dar lugar a falsos negativos.
- Se puede detectar un bajo número de copias del RNA molde diana por debajo del límite de detección, pero los resultados pueden no ser reproducibles.
- Existe la posibilidad de falsos positivos debido a la contaminación cruzada con virus Chikungunya, ya sea por muestras que contienen altas concentraciones de RNA molde diana o por contaminación por arrastre a partir de productos de PCR de reacciones anteriores.

Adjunto 1: Compatibilidad de los termocicladores a tiempo real más usuales

Los termocicladores más usuales se enumeran en la siguiente tabla separados por tipo de tubo. Si no encuentra su termociclador en la siguiente lista, póngase en contacto con su proveedor

Termocicladores con bloque de bajo perfil	Termocicladores con bloque de alto perfil
Applied Biosystems	Applied Biosystems
7500 Fast Real-Time PCR System	7300/7500/7900HT Real-Time PCR
7500 Fast Dx Real-Time PCR System	ABI PRISM 7000
QuantStudio™ 12K Flex 96-well Fast	ABI PRISM 7700
QuantStudio™ 6 Flex 96-well Fast	QuantStudio™ 12K Flex 96-well
QuantStudio™ 7 Flex 96-well Fast	QuantStudio™ 6 Flex 96-well
QuantStudio™ 3 Real-Time PCR System	QuantStudio™ 7 Flex 96-well
QuantStudio™ 5 Real-Time PCR System	QuantStudio™ 3 Real-Time PCR
StepOne Plus™ Real-Time PCR System	QuantStudio™ 5 Real-Time PCR
StepOne™ Real-Time PCR System	ViiA™ 7 Real-Time PCR
ViiA™ 7 Fast Real-Time PCR System	Bio-Rad
Bio-Rad	CFX96 Touch™ Deep Well Real-Time PCR Detection
CFX96 Touch™ Real-Time PCR Detection System	iCycler iQ™ Real-Time PCR
Mini Opticon™ Real-Time PCR Detection System	iCycler iQ™5 Real-Time PCR
Roche	MyiQ™ Real-Time PCR
LightCycler®480 Real-Time PCR System	MyiQ™2 Real-Time PCR
LightCycler®96 Real-Time PCR System	Eppendorf
Agilent Technologies	Mastercycler™ep <i>realplex</i>
AriaMx Real-Time PCR System	Stratagene / Agilent Technologies
Qiagen	Mx3000P™ Real Time PCR System
Rotor-Gene®	Mx3005P™ Real Time PCR System
Cepheid	Analytik Jena Biometra
SmartCycler®	TOptical
	qTOWER 2.0
	Abbott
	Abbott m2000 RealTime System
	BIONEER
	Exicycler™ 96
	DNA-Technology
	DTlite Real-Time PCR System*
	DTprime Real-time*
	Qiagen
	Rotor-Gene®
	Cepheid
	SmartCycler®

* Ver Adjunto III para configurar los valores de exposición.

Adjunto III: Configuración valores de exposición

Los valores de exposición de algunos termocicladores se deben ajustar para su correcto funcionamiento. Establecer los valores de exposición de la siguiente manera:

Termociclador	Canal Vitassay	Valor de exposición
DTlite Real-Time PCR System (DNA-Technology)	FAM	150
	HEX	500
	ROX	500
	Cy5	100
DTprime Real-time Detection Thermal Cycler (DNA- Technology)	FAM	150
	HEX	3000
	ROX	2000
	Cy5	1500

Intended use

Vitassay qPCR Chikungunya allows the qualitative detection of Chikungunya virus by real-time RT-PCR in clinical samples. The product is intended for use in the diagnosis of Chikungunya virus infections alongside clinical data of the patient and other laboratory tests outcomes.

References

Vitassay qPCR Chikungunya 4 x 8-well strip, low profile 7041001
 Vitassay qPCR Chikungunya 4 x 8-well strip, high profile 7042001

Materials/reagents provided

Código	Reactivo/Material	Color	Cantidad
7041S001/ 7042S001	Chikungunya Virus strips high/low profile	-	4x8-well strip
7C001	Chikungunya Virus Positive Control	red	1 vial
7001A	PCR grade water	white	1 vial x 1 mL
7002B	Rehydration Buffer	green	1 vial x 1 mL
7003N	Negative control	yellow	1 vial x 1 mL
7004O	Optical caps	-	4x8 cap strip

Transport and storage

- The reagents and the test can be shipped and stored at 2-40°C until expiration date stated in the label.
- The resuspended positive control should be stored at -20°C. In order to avoid repeated freeze/thaw cycles, we recommend to separate in aliquots.
- Keep all reagents of in the dark.

Additional equipment and material required

- Real-time PCR instrument (thermocycler) (Attached I)
- RNA extraction kit
- Centrifuge for 1.5 mL tubes
- Vortex
- Micropipettes (1-20 µL, 20-200 µL)
- Filter tips
- Powder-free disposal gloves

Summary

Chikungunya virus (CHIKV) is a mosquito-borne virus (arbovirus) first described during an outbreak in southern Tanzania in 1952. It's transmitted to humans mainly by infected mosquitos of the *Aedes genus*, specifically *Ae. aegypti* and *Ae. albopictus*. Although perinatal vertical transmission has also been reported.

Since 1952, CHIKV has caused numerous well-documented outbreaks and epidemics in Africa, Asia, Europe, the South Pacific and Americas, involving millions of people, and often separated by periods of more than 10 years.

After the bite of an infected mosquito, onset of illness occurs usually between 4 and 8 days but can range from 2 to 12 days. Chikungunya infection is characterised by an abrupt onset of chills, fever reaching up to 40 °C, vomiting, nausea, headache, arthralgia (joint pain), and in some cases, maculopapular rash. Severe joint and muscular pain is the main and the most problematic symptom of chikungunya. The pain is so intense that makes movement very difficult and prostrates its victims.

Laboratory diagnosis is generally accomplished by testing serum or plasma to detect virus, viral nucleic acid, or virus-specific immunoglobulin (Ig) M and neutralizing antibodies. Viral culture may detect virus in the first 3 days of illness; however, chikungunya virus should be handled under biosafety level (BSL) 3 conditions. During the first 8 days of illness, chikungunya viral RNA can often be identified in serum. Chikungunya virus antibodies normally develop toward the end of the first week of illness. Therefore, to definitively rule out the diagnosis, convalescent-phase samples should be obtained from patients whose acute-phase samples test negative.

Real Time RT-PCR is a common detection method due to cross-reactivity of Chikungunya antibodies with other arbovirus limits the use of serology.

Principle of the test

Vitassay qPCR Chikungunya is designed for the identification of Chikungunya virus in clinical specimens to aid in the assessment of infections caused by this virus.

Vitassay qPCR Chikungunya is based on the real-time amplification of specific conserved fragments of the *NSP1* gene encoded by the Chikungunya virus genome. The viral RNA extracted is transcribed into cDNA using a specific primer by reverse transcription step followed immediately in the same well by polymerase chain reaction. The presence of Chikungunya virus is detected by an increase in observed fluorescence during the reaction upon hydrolysis of the fluorescent probe.

Vitassay qPCR Chikungunya is a ready-to used test which contains in each well all the necessary reagents for real-time PCR assay in a stabilized format. In addition, an internal control allows the detection of a possible reaction inhibition. The amplification of the target sequence is detected through the FAM channel whereas the internal control (IC) in HEX, VIC or JOE channel (depending on the equipment used).

Precautions

- For *in vitro* diagnostic use.
- Do not use after expiration date.
- Do not mix components from different kits and/or lots.
- Do not use if package is open or damaged.
- Follow Good Laboratory Practices. Wear protective clothing, use disposal gloves, goggles and mask.
- Do not eat, drink or smoke in the laboratories.
- The laboratory process must be one-directional, it should begin in the Extraction Area and then move to the Amplification and Detection Areas. Do not return samples, equipment and reagents to the area in which the previous step was performed.
- Specimens and reagents/materials that have been exposed to them must be treated as potentially infectious. Take necessary precautions during the collection, storage, treatment and disposal of samples.
- Regular decontamination of commonly used equipment is recommended, especially micropipettes and work surfaces.

Procedures

Specimen collection, processing and RNA extraction

For pretreatment and nucleic acid isolation, it is recommended to use your optimized manual or automatic system compatible with real-time PCR technology. The assay has been validated with the following extraction kits:

Maxwell® 16 Viral Total Nucleic Acid Purification Kit (Promega).

QIAamp Viral RNA Mini Kit, QIAcube instrument (Qiagen).

Invisorb® Spin Universal Kit (Strattec).

Positive control preparation

Reconstitute the lyophilized Chikungunya Virus Positive Control (red tube) in the 100 μ L of PCR grade water (white tube) supplied. To ensure a complete resuspension, vortex the tube thoroughly. After first use, dispense into aliquots in order to avoid multiple freeze-thaw cycles, and store them at -20°C.

This component contains high copies number template and is a very significant contamination risk. Therefore, we recommend open and manipulate it in a separate laboratory area away from the other components.

Reaction setup

- Separate the number of required reactions including samples and controls. Remember that one positive and one negative controls must be included in each run. Peel off protective aluminium seal from the strips.
- Pipette 15 μ L of resuspension buffer (green tube) into each well.
- Pipette 5 μ L of RNA sample, negative and positive controls into each well
- Cover the wells with the caps provided. Spin down briefly (optional).
- Place the strips in the Real-time PCR instrument.

Programm your thermocycler.

Set your thermocycler following the conditions below:

Step	Temperature	Time	Cycles
Reverse transcription	45°C	15 min	1
Initial denaturation	95°C	2 min	1
Denaturation	95°C	10 sec	45
Annealing/Extension (Data collection*)	60°C	50 sec	

Set the fluorescence data collection during the extension step (*) through the FAM (Chikungunya virus) and HEX, JOE or VIC channels (Internal Control (IC)). If you use

the Applied Biosystems 7500 Fast, the Applied Biosystems StepOne™ or the Stratagene Mx3005P™ check that passive reference option ROX is none (attached II).

Analysis and interpretation of results

The analysis of the results is done by the software itself of the used real-time PCR system following manufacturer’s instructions.

For a valid diagnostic test run, the following control conditions must be met:

Positive control

The positive controls used in each run, must show an amplification curve for Chikungunya virus, which validates the reaction.

Negative control

The negative controls included in each run, must show the absence of signal for Chikungunya virus, which validates the reaction.

The experiment seems to be fail if there is signal of amplification in negative control or absence of signal in the positive well. The assay should be repeated.

The result interpretation is summarized in the following table:

Virus Chikungunya FAM	Control Interno HEX	Negative Control	Positive Control	Interpretation
+	+/-	-	+	Virus Chikungunya Positive
-	+	-	+	Virus Chikungunya Negative
+	+	+	+	Invalid
-	-	-	-	Invalid

+ Positive: Amplification signal

- Negative: No amplification signal

If the negative samples not show a positive result for the internal control, they should be retested from the diluted original sample 1:10 or the nucleic acid extraction has to be repeated due to possible problems caused by PCR inhibitors.

Quality Control

In order to confirm the appropriate performance of the molecular diagnostic technique, an Internal Control (IC) is included in each reaction. Besides, a positive and a negative control must be included in each assay to interpret the results correctly.

Performance evaluation

Clinical sensitivity and specificity

Overall, 10 clinical samples were tested by Real-time PCR using: Vitassay qPCR Chikungunya and RealStar® Chikungunya Virus RT-PCR Kit (Altona Diagnostics). Chikungunya virus was detected in all samples by Vitassay qPCR Chikungunya. The results were concordant between all test. Vitassay qPCR Chikungunya shows a high sensitivity and specificity to detect Chikungunya virus.

Analytical sensitivity

The analytical sensitivity was determined by analysis of 10-fold dilution series of Chikungunya virus template ranging from 10^7 to 10^1 copies/rxn. This assay has a detection limit of ≥ 10 viral RNA copies per reaction.

Analytical specificity

The analytical specificity for Chikungunya virus was tested within the panel of following microorganisms, where no cross-reactivity was seen between any of the species:

Chikungunya Virus		
Zika virus strain MR 766	Dengue 4 virus strain H241	West Nile virus Heja
Dengue 1 virus strain Hawaii	St Louis Encephalitis virus strain 17D	West Nile virus Ug37
Dengue 2 virus strain New Guinea C	West Nile virus strain H160/99	Yellow Fever virus strain 17D
Dengue 3 virus strain H87		

Analytical reactivity

The reactivity of Vitassay qPCR Chikungunya was confirmed by the real-time amplification using Chikungunya virus strain S27 Petersfield as template.

Compatibles real-time PCR equipment

Vitassay qPCR Chikungunya has been validated on the following equipments:

- LightCycler Z480, 480II (Roche)
- Cobas Z480 (Roche)
- 7500 Fast Real-Time PCR System (Applied Biosystems)
- StepOne™ Real-Time PCR System (Applied Biosystems)
- CFX96 Touch™ Real-Time PCR Detection System (Bio-Rad)
- AriaMx Real-Time PCR System (Agilent Technologies)
- DTPPrime Real Time Detection Thermal Cycler (DNA-Technology)
- DTLite Real Time PCR System (DNA-Technology)
- Rotor-Gene® Q (Qiagen)
- SmartCycler® (Cepheid)

For Rotor-Gene® Q and SmartCycler® thermocyclers the product should be reconstituted following the procedure and transferred into specific Rotor-Gene® Q and/or SmartCycler tubes.

When using the Applied Biosystems 7500 Fast with strips it is recommend to place a plate holder for the tubes (Ref. PN 4388506).

Limitations

- This test provides a presumptive diagnosis of Chikungunya virus infection. All results must be interpreted together with other clinical information and laboratory findings available to the physician.
- This assay was tried with samples in transport medium. The use of other samples has not been established.
- The quality of the test depends on the quality of the sample; proper RNA from clinical specimens must be extracted. Unsuitable collection, storage and/or transport of specimens may give false negative results.
- Extremely low levels of target below the limit of detection may be detected, but results may not be reproducible.
- There is a possibility of false positive results due to cross-contamination by Chikungunya virus, either samples containing high concentrations of target RNA or contamination due to PCR products from previous reactions.

Attached I: Compatibility of the most common real-time PCR equipment

The most common thermocyclers are listed in the following table separated by tube type. If you do not find your thermocycler in the list below, please contact with your supplier.

Low profile Block Thermocyclers
Applied Biosystems
7500 Fast Real-Time PCR System
7500 Fast Dx Real-Time PCR System
QuantStudio™ 12K Flex 96-well Fast
QuantStudio™ 6 Flex 96-well Fast
QuantStudio™ 7 Flex 96-well Fast
QuantStudio™ 3 Real-Time PCR System
QuantStudio™ 5 Real-Time PCR System
StepOne Plus™ Real-Time PCR System
StepOne™ Real-Time PCR System
ViiA™ 7 Fast Real-Time PCR System
Bio-Rad
CFX96 Touch™ Real-Time PCR Detection System
Mini Opticon™ Real-Time PCR Detection System
Roche
LightCycler® 480 Real-Time PCR System
LightCycler® 96 Real-Time PCR System
Agilent Technologies
AriaMx Real-Time PCR System
Qiagen
Rotor-Gene®
Cepheid
SmartCycler®

High profile Block Thermocyclers
Applied Biosystems
7300/7500/7900HT Real-Time PCR
ABI PRISM 7000
ABI PRISM 7700
QuantStudio™ 12K Flex 96-well
QuantStudio™ 6 Flex 96-well
QuantStudio™ 7 Flex 96-well
QuantStudio™ 3 Real-Time PCR
QuantStudio™ 5 Real-Time PCR
ViiA™ 7 Real-Time PCR
Bio-Rad
CFX96 Touch™ Deep Well Real-Time PCR Detection
iCycler iQ™ Real-Time PCR
iCycler iQ™5 Real-Time PCR
MyiQ™ Real-Time PCR
MyiQ™2 Real-Time PCR
Eppendorf
Mastercycler™ep <i>realplex</i>
Stratagene / Agilent Technologies
Mx3000P™ Real Time PCR System
Mx3005P™ Real Time PCR System
Analytik Jena Biometra
TOptical
qTOWER 2.0
Abbott
Abbott m2000 RealTime System
BIONEER
Exicycler™ 96
DNA-Technology
DTlite Real-Time PCR System*
DTprime Real-time*
Qiagen
Rotor-Gene®
Cepheid
SmartCycler®

* See Attached III to configure exposure settings.

Attached III: Optical measurement exposure setting

The exposure values of some thermocyclers must be adjusted for proper operation. Set exposure values as follows:

Thermocycler	Vitassay channel	Exposure values
DTlite Real-Time PCR System (DNA-Technology)	FAM	150
	HEX	500
	ROX	500
	Cy5	100
DTprime Real-time Detection Thermal Cycler (DNA- Technology)	FAM	150
	HEX	3000
	ROX	2000
	Cy5	1500

Bibliography/Bibliografía

1. Genetic and serologic properties of Chikungunya virus associated with an epidemic, Yap State, Micronesia, 2007. Lanciotti RS, Kosoy OL, Laven JJ, Velez JO, Lambert AJ, Johnson AJ, Stanfield SM, Duffy MR. Emerg Infect Dis. 2008 Aug;14(8):1232-9.
2. Quantitative real-time PCR detection of Chikungunya virus and evaluation with field-caught mosquitoes. Faye O, Faye O, Diallo D, Diallo M, Weidmann M, Sall AA. Virol J. 2013 Oct 22;10:311.
3. Detection of Chikungunya virus in urine. Gourinat. AC, O'Connor O, Calvez E, Goarant C, Dupont-Rouzeyrol M. Emerg Infect Dis. 2015 Jan;21(1):84-6.
4. Centers for Disease Control and Prevention. Chikungunya Virus (<http://www.cdc.gov/Chikungunya/index.html>).
5. World Health Organization. Chikungunya virus and potential complications (<http://www.who.int/emergencies/Chikungunya-virus/en/>).

Trademarks

CFX™ and IQ5™ are registered trademarks of Bio-Rad Laboratories.

ABI®, QuantStudio™, StepOnePlus™ and ViiA™ are registered trademarks of Thermo Fisher Scientific Inc.

LightCycler® is a registered trademark of Roche.










Mx3000P™ and Mx3005™ are registered trademarks of Agilent Technologies.

Mastercycler™ is a registered trademark of Eppendorf.

Rotor-Gene® Q is a registered trademark of Qiagen.

SmartCycler® is a registered trademark of Cepheid

Symbols for IVD components and reagents/ Símbolos utilizados para componentes y reactivos IVD

	Producto para diagnóstico <i>in vitro</i> For <i>in vitro</i> diagnostic use only		Almacenar en lugar seco Keep dry
	Consultar las instrucciones de uso Consult instructions for use		Limitación de temperatura Temperature limitation
	Fecha de caducidad Use by		Fabricante Manufacturer
	Número de lote Lot number		Contiene <n> test Contains sufficient for <n> test
DIL	Diluyente de muestra Buffer Sample diluent		Número de referencia Catalogue number



Vitassay Healthcare S.L.U. Parque Tecnológico Walqa
Ctra. N-330 Km. 566 • 22197 Huesca (Spain)

www.vitassay.com